

**MEXPO INTERNATIONAL INC.**

2828 Faber Street  
Union City, CA 94587-1204

**Declaration of Conformity  
(Face Mask)**

**PRODUCT DESCRIPTION**

- 1. Product Name: Face Mask (non-woven products)
- 2. Product Classification: Class I under **Medical Device Regulations (EU) 2017/745 Annex VIII Rule 1**

**ADDRESS:**

**Mexpo International Inc.**  
 2828 Faber Street  
 Union City, CA 94587-1204, USA  
 Tel : +1 (510) 489-6800  
 Fax : +1 (510) 489-3111  
 E-mail : [blossomglo@aol.com](mailto:blossomglo@aol.com)  
 E-mail: [tim@mexpo-glove.com](mailto:tim@mexpo-glove.com)

**EC Representative:**  
**Blossom Europe, S. L**  
 Paseo de Recoletos 37-41,  
 28004 Madrid, Spain

**Brand Owner**  
 white-med GmbH  
 Marburger Straße 251  
 35396 Gießen, Germany

We, **MEXPO International, Inc.** as the legal manufacturer declare under our sole responsibility that the medical devices listed below conform to the requirement of the **Medical Device Regulations (EU) 2017/745**.

**Face Mask Product:**

**1) N700 Sensitive New 3 PLY Latex Free Earloop Face Mask**

COLOR:	ITEM #	UDI #:
Blue	N700BG	4046144024003
Green	N700GG	4046144024027
Lila	N700LG	4046144024065
Pink	N700PG	4046144024034
White	N700WG	4046144024041

**2) N770 Peppler Care 3 PLY Latex Free Earloop Face Mask**

COLOR:	ITEM #	UDI #:
Blue	N770BG	4046144011881
Green	N770GG	4046144011898
Lila	N770LG	4046144011911
Pink	N770PG	4046144011928
White	N770WG	4046144011935

**Basic UDI-DI: 72386035177RA**

It is declared that above devices meet the requirement of the **Medical Device Regulation EU/2017/745**.


We further declare that Blossom Europe, SL, located at Paseo de Recoletos 37-41, 28004 Madrid, SPAIN, is an European Authorized Representative.

PPE Regulations (EU) 2016/425- Cat III and, where such is the case, with the following standards.  
DIN EN 14683, ASTM F2299/F2299M-03(2017), ASTM F2101-19, ASTM F2100-20, ASTM F1862/F1862M-17, ASTM F1494-14.

In accordance with Annex VIII, **Medical Device Regulation EU/2017/745**, is listed above are non-invasive transient devices and are Class I devices under Rule 1 as Rules 2, 3, and 4 do not apply.

The following person is responsible for signing of this Document:

Name and Address: Mexpo International, Inc., 2828 Faber Street, Union City, California 94587-1204, USA

Authorized Signature: 

Date: September 30, 2021

Name of responsible Person: TIM THAI

Position: President